



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

(Certificate No : L0954-250718)

This is to certify that

Gong-Hung Technology Co., Ltd.

GCH Technology Calibration Laboratory(Length)

No.1, Ln. 36, Wenfeng St., Fengshan Dist., Kaohsiung City 830, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2017 ; CNS 17025:2018

Accreditation Number : 0954

Originally Accredited : January 01, 2003

Effective Period : July 20, 2025 to July 19, 2028

Accredited Scope : Calibration Field, see described in the Appendix



Scan to verify

Yi-Ling Chen

Yi-Ling Chen
President, Taiwan Accreditation Foundation
July 18, 2025

Accreditation Number : 0954

Laboratory Head : HUNG, Chuan-Hsi

Length

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA2003 vernier caliper digimatic caliper dial caliper	Mitutoyo /515-556-2 KOBA /600 mm	In-house method: caliper calibration procedure (Document No.: GCH-SCP-L01)	0	mm	300	mm	vernier outside diameter (resolution: 0.02 mm)	0.02	mm
			0	mm	300	mm	vernier inside diameter (resolution: 0.02 mm)	0.02	mm
			0	mm	300	mm	vernier outside diameter (resolution: 0.05 mm)	0.05	mm
			>300	mm	600	mm	vernier outside diameter (resolution: 0.05 mm)	0.05	mm
			0	mm	300	mm	digimatic inside diameter (resolution: 0.01 mm)	0.05	mm
			>300	mm	600	mm	digimatic inside diameter (resolution: 0.01 mm)	0.05	mm
			0	mm	200	mm	digimatic outside diameter (resolution: 0.01 mm)	0.02	mm
			0	mm	200	mm	digimatic inside diameter (resolution: 0.01 mm)	0.02	mm
			0	mm	300	mm	digimatic outside diameter (resolution: 0.01 mm)	0.02	mm
			>300	mm	600	mm	digimatic outside diameter (resolution: 0.01 mm)	0.02	mm
			0	mm	300	mm	digimatic inside diameter (resolution: 0.01 mm)	0.02	mm
			>300	mm	600	mm	digimatic inside diameter (resolution: 0.01 mm)	0.02	mm



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA2003 vernier caliper digimatic caliper dial caliper	Mitutoyo /515-556-2 KOBA /600 mm	In-house method: caliper calibration procedure (Document No.: GCH-SCP-L01)	0	mm	200	mm	dial outside diameter (resolution: 0.01 mm)	0.02	mm
			0	mm	200	mm	dial inside diameter (resolution: 0.01 mm)	0.02	mm
			0	mm	300	mm	dial outside diameter (resolution: 0.02 mm)	0.03	mm
			0	mm	300	mm	dial inside diameter (resolution: 0.02 mm)	0.03	mm

Approval Signatory: HUNG, Chuan-Hsi; CHEN, Chih-Chuan

KA2005 digimatic outside micrometer vernier outside micrometer digit outside micrometer	Mitutoyo /BM1-10N-0 /PD Mitutoyo /BM1-8M-0 /D Mitutoyo /30 mm-90 mm	In-house method: outside micrometer calibration procedure (Document No.: GCH-SCP-L03)	0	mm	25	mm	digimatic (resolution: 0.001 mm)	0.003	mm
			25	mm	50	mm	digimatic (resolution: 0.001 mm)	0.003	mm
			50	mm	75	mm	digimatic (resolution: 0.001 mm)	0.003	mm
			75	mm	100	mm	digimatic (resolution: 0.001 mm)	0.003	mm
			0	mm	25	mm	vernier (resolution: 0.01 mm)	0.01	mm
			25	mm	50	mm	vernier (resolution: 0.01 mm)	0.01	mm
			50	mm	75	mm	vernier (resolution: 0.01 mm)	0.01	mm
			75	mm	100	mm	vernier (resolution: 0.01 mm)	0.01	mm
			0	mm	25	mm	vernier (resolution: 0.001 mm)	0.003	mm
			25	mm	50	mm	vernier (resolution: 0.001 mm)	0.003	mm
			50	mm	75	mm	vernier (resolution: 0.001 mm)	0.003	mm
			75	mm	100	mm	vernier (resolution: 0.001 mm)	0.003	mm
			0	mm	25	mm	digit (resolution: 0.001 mm)	0.003	mm

Approval Signatory: HUNG, Chuan-Hsi; CHEN, Chih-Chuan



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA2008 2D Height Gauge (on-site calibration included)	Step Gauge	In-house method: Calibration Procedure for 2D Height Gauge (Document No.: GCH-SCP-L06) (on-site calibration included)	0	mm	600	mm	resolution 0.0001 mm	11	μm
			0	mm	600	mm	resolution 0.001 mm	11	μm
			0	mm	600	mm	resolution 0.0001 mm (on- site calibration)	14	μm
			0	mm	600	mm	resolution 0.001 mm (on- site calibration)	14	μm

Approval Signatory: HUNG, Chuan-Hsi; CHEN, Chih-Chuan

KA2008 Height Gauge (on-site calibration included)	Step Gauge Mitutoyo /515-743	In-house method: Calibration Procedure for Height Gauge (on-site calibration included) (Document No.: GCH-SCP-L05)	0	mm	600	mm	digimatic (resolution: 0.01 mm) (on-site calibration included)	0.02	mm
			0	mm	600	mm	dial (resolution: 0.01 mm) (on-site calibration included)	0.02	mm

Approval Signatory: HUNG, Chuan-Hsi; CHEN, Chih-Chuan

KA2010 dial indicator digimatic indicator	Mitutoyo /1.005 mm-9 mm Mitutoyo /BMI-10N-0/PD Mitutoyo /516-115-10 Mitutoyo /30 mm-90 mm Mitutoyo/170-102	In-house method: indicator calibration procedure (Document No.: GCH-SCP-L02)	0	mm	5	mm	dial (resolution: 0.001 mm)	0.004	mm
			0	mm	20	mm	dial (resolution: 0.01 mm)	0.01	mm
			0	mm	50	mm	dial (resolution: 0.01 mm)	0.01	mm
			0	mm	25	mm	digimatic (resolution: 0.01 mm)	0.01	mm
			0	mm	25	mm	digimatic (resolution: 0.001 mm)	0.004	mm
			0	mm	50	mm	digimatic (resolution: 0.01 mm)	0.01	mm
			0	mm	50	mm	digimatic (resolution: 0.001 mm)	0.001	mm

Approval Signatory: HUNG, Chuan-Hsi; CHEN, Chih-Chuan



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA2010 Test indicat	Dial Gauge Calibrator	In-house method: Test indicat Calibration Procedure (Document No.: GCH-SCP-L04)	0	mm	0.2	mm	resolution: 0.001 mm	0.004	mm
			0	mm	0.2	mm	resolution: 0.002 mm	0.004	mm
			0	mm	0.8	mm	resolution: 0.01 mm	0.01	mm
Approval Signatory: HUNG, Chuan-Hsi; CHEN, Chih-Chuan									
KA4008 3D Coordinate Measuring Machine (on-site calibration)	Mitutoyo /30 mm Mitutoyo /BM1-8M-0 KOBA/ (200 to 500) mm KOBA /600 mm KOBA /700 mm SAGER /24.99505 mm	In-house method: Calibration Procedure for 3D Coordinate Measuring Machine (on-site calibration) (Document No.: GCH-SCP-L08)	0	mm	500	mm	X Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	$[(2.4)^2 + (9.1 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	Y Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	$[(3.5)^2 + (9.1 L)^2]^{1/2}$, Length L in m	μm
			0	mm	500	mm	Z Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	$[(2.2)^2 + (9.1 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	space/ (20 ± 2) °C / (resolution: 0.001 mm)	$[(3.6)^2 + (9.1 L)^2]^{1/2}$, Length L in m	μm
							probing error/ (20 ± 2) °C / (resolution: 0.001 mm)	1	μm



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA4008 3D Coordinate Measuring Machine (on-site calibration)	Mitutoyo /30 mm Mitutoyo /BM1-8M-0 KOB A/ (200 to 500) mm KOB A/ /600 mm KOB A/ /700 mm SAGER /24.99505 mm	In-house method: Calibration Procedure for 3D Coordinate Measuring Machine (on-site calibration) (Document No.: GCH-SCP-L08)	0	mm	500	mm	X Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	$[(2.4)^2 + (22 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	Y Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	$[(3.5)^2 + (22 L)^2]^{1/2}$, Length L in m	μm
			0	mm	500	mm	Z Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	$[(2.2)^2 + (22 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	space/ (20 ± 5) °C / (resolution: 0.001 mm)	$[(3.6)^2 + (22 L)^2]^{1/2}$, Length L in m	μm
							probing error / (20 ± 5) °C / (resolution: 0.001 mm)	1	μm
			0	mm	500	mm	X Axis/ (20 ± 10) °C / (resolution: 0.001 mm)	$[(2.4)^2 + (43 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	Y Axis/ (20 ± 10) °C / (resolution: 0.001 mm)	$[(3.5)^2 + (43 L)^2]^{1/2}$, Length L in m	μm



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA4008 3D Coordinate Measuring Machine (on-site calibration)	Mitutoyo /30 mm Mitutoyo /BM1-8M-0 KOB A/ (200 to 500) mm KOB A/ /600 mm KOB A/ /700 mm SAGER /24.99505 mm	In-house method: Calibration Procedure for 3D Coordinate Measuring Machine (on-site calibration) (Document No.: GCH-SCP-L08)	0	mm	500	mm	Z Axis/ (20 ± 10) °C / (resolution: 0.001 mm)	$[(2.2)^2 + (43 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	space/ (20 ± 10) °C / (resolution: 0.001 mm)	$[(3.6)^2 + (43 L)^2]^{1/2}$, Length L in m	μm
							probing error / (20 ± 10) °C / (resolution: 0.001 mm)	2	μm
			0	mm	500	mm	X Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	$[(2.4)^2 + (9.1 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	Y Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	$[(2.8)^2 + (9.1 L)^2]^{1/2}$, Length L in m	μm
			0	mm	500	mm	Z Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	$[(2.0)^2 + (9.1 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	space/ (20 ± 2) °C / (resolution: 0.0001 mm)	$[(3.4)^2 + (9.1 L)^2]^{1/2}$, Length L in m	μm



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA4008 3D Coordinate Measuring Machine (on-site calibration)	Mitutoyo /30 mm Mitutoyo /BM1-8M-0 KOB A/ (200 to 500) mm KOB A/ /600 mm KOB A/ /700 mm SAGER /24.99505 mm	In-house method: Calibration Procedure for 3D Coordinate Measuring Machine (on-site calibration) (Document No.: GCH-SCP-L08)					probing error / (20 ± 2) °C / (resolution: 0.0001 mm)	0.4	μm
			0	mm	500	mm	X Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	$[(2.4)^2 + (22 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	Y Axis/ (20 ± 5) °C/ (resolution: 0.0001 mm)	$[(2.8)^2 + (22 L)^2]^{1/2}$, Length L in m	μm
			0	mm	500	mm	Z Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	$[(2.0)^2 + (22 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	space/ (20 ± 5) °C / (resolution: 0.0001 mm)	$[(3.4)^2 + (22 L)^2]^{1/2}$, Length L in m	μm
							probing error / (20 ± 5) °C / (resolution: 0.0001 mm)	0.6	μm
			0	mm	500	mm	X Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	$[(2.4)^2 + (43 L)^2]^{1/2}$, Length L in m	μm



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA4008 3D Coordinate Measuring Machine (on-site calibration)	Mitutoyo /30 mm Mitutoyo /BM1-8M-0 KOBA/ (200 to 500) mm KOBA /600 mm KOBA /700 mm SAGER /24.99505 mm	In-house method: Calibration Procedure for 3D Coordinate Measuring Machine (on-site calibration) (Document No.: GCH-SCP-L08)	0	mm	700	mm	Y Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	$[(2.8)^2 + (43 L)^2]^{1/2}$, Length L in m	μm
			0	mm	500	mm	Z Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	$[(2.0)^2 + (43 L)^2]^{1/2}$, Length L in m	μm
			0	mm	700	mm	space/ (20 ± 10) °C / (resolution: 0.0001 mm)	$[(3.4)^2 + (43 L)^2]^{1/2}$, Length L in m	μm
							probing error / (20 ± 10) °C / (resolution: 0.0001 mm)	0.9	μm
Approval Signatory: HUNG, Chuan-Hsi; CHEN, Chih-Chuan									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA4010 video measurement system (linearscale type) (on-site calibration)	Mitutoyo /BM1-8M-0 Nikon/50 mm Nikon/300 mm Mitutoyo /182-532 Mitutoyo /HL3-1000	In-house method: Calibration Procedure for video measurement system (on-site calibration) (Document No.: GCH-SCP-L07)	0	mm	50	mm	X Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	2.3	µm
			0	mm	300	mm	X Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	2.4	µm
			0	mm	500	mm	X Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	2.6	µm
			0	mm	780	mm	X Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	3.3	µm
			0	mm	50	mm	Y Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	1.4	µm
			0	mm	300	mm	Y Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	1.7	µm
			0	mm	500	mm	Y Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	2.1	µm
			0	mm	580	mm	Y Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	2.4	µm
			0	mm	150	mm	Z Axis/ (20 ± 2) °C / (resolution: 0.001 mm)	2.0	µm
			0	mm	50	mm	Exy diagonal axis/ (20 ± 2) °C / (resolution: 0.001 mm)	3.9	µm
			0	mm	300	mm	Exy diagonal axis/ (20 ± 2) °C / (resolution: 0.001 mm)	3.7	µm
			0	mm	500	mm	Exy diagonal axis/ (20 ± 2) °C / (resolution: 0.001 mm)	3.7	µm
			0	mm	900	mm	Exy diagonal axis/ (20 ± 2) °C / (resolution: 0.001 mm)	4.2	µm
			0	mm	50	mm	X Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	2.3	µm
			0	mm	300	mm	X Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	2.8	µm
			0	mm	500	mm	X Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	3.8	µm



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA4010 video measurement system (linearscale type) (on-site calibration)	Mitutoyo /BM1-8M-0 Nikon/50 mm Nikon/300 mm Mitutoyo /182-532 Mitutoyo /HL3-1000	In-house method: Calibration Procedure for video measurement system (on-site calibration) (Document No.: GCH-SCP-L07)	0	mm	780	mm	X Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	5.4	µm
			0	mm	50	mm	Y Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	1.4	µm
			0	mm	300	mm	Y Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	2.4	µm
			0	mm	500	mm	Y Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	3.5	µm
			0	mm	580	mm	Y Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	4.0	µm
			0	mm	150	mm	Z Axis/ (20 ± 5) °C / (resolution: 0.001 mm)	3.5	µm
			0	mm	50	mm	Exy diagonal axis/ (20 ± 5) °C / (resolution: 0.001 mm)	4.0	µm
			0	mm	300	mm	Exy diagonal axis/ (20 ± 5) °C / (resolution: 0.001 mm)	3.7	µm
			0	mm	500	mm	Exy diagonal axis/ (20 ± 5) °C / (resolution: 0.001 mm)	4.3	µm
			0	mm	900	mm	Exy diagonal axis/ (20 ± 5) °C / (resolution: 0.001 mm)	6.5	µm
			0	mm	50	mm	X Axis/ (20 ± 10) °C/ (resolution: 0.001 mm)	2.3	µm
			0	mm	300	mm	X Axis/ (20 ± 10) °C/ (resolution: 0.001 mm)	4.2	µm
			0	mm	500	mm	X Axis/ (20 ± 10) °C/ (resolution: 0.001 mm)	6.6	µm
			0	mm	780	mm	X Axis/ (20 ± 10) °C/ (resolution: 0.001 mm)	11	µm
			0	mm	50	mm	Y Axis/ (20 ± 10) °C/ (resolution: 0.001 mm)	1.5	µm
			0	mm	300	mm	Y Axis/ (20 ± 10) °C/ (resolution: 0.001 mm)	4.0	µm



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA4010 video measurement system (linearscale type) (on-site calibration)	Mitutoyo /BM1-8M-0 Nikon/50 mm Nikon/300 mm Mitutoyo /182-532 Mitutoyo /HL3-1000	In-house method: Calibration Procedure for video measurement system (on-site calibration) (Document No.: GCH-SCP-L07)	0	mm	500	mm	Y Axis/ (20 ± 10) °C/ (resolution: 0.001 mm)	6.5	µm
			0	mm	580	mm	Y Axis/ (20 ± 10) °C/ (resolution: 0.001 mm)	7.5	µm
			0	mm	150	mm	Z Axis/ (20 ± 10) °C/ (resolution: 0.001 mm)	6.5	µm
			0	mm	50	mm	Exy diagonal axis/ (20 ± 10) °C / (resolution: 0.001 mm)	4.0	µm
			0	mm	300	mm	Exy diagonal axis/ (20 ± 10) °C / (resolution: 0.001 mm)	4.7	µm
			0	mm	500	mm	Exy diagonal axis/ (20 ± 10) °C / (resolution: 0.001 mm)	6.8	µm
			0	mm	900	mm	Exy diagonal axis/ (20 ± 10) °C / (resolution: 0.001 mm)	12	µm
			0	mm	50	mm	X Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	2.8	µm
			0	mm	300	mm	X Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	2.5	µm
			0	mm	500	mm	X Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	2.7	µm
			0	mm	780	mm	X Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	3.3	µm
			0	mm	50	mm	Y Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	1.7	µm
			0	mm	300	mm	Y Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	1.9	µm
			0	mm	500	mm	Y Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	2.3	µm
			0	mm	580	mm	Y Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	2.5	µm
			0	mm	150	mm	Z Axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	2.0	µm



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA4010 video measurement system (linearscale type) (on-site calibration)	Mitutoyo /BM1-8M-0 Nikon/50 mm Nikon/300 mm Mitutoyo /182-532 Mitutoyo /HL3-1000	In-house method: Calibration Procedure for video measurement system (on-site calibration) (Document No.: GCH-SCP-L07)	0	mm	50	mm	Exy diagonal axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	3.3	µm
			0	mm	300	mm	Exy diagonal axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	3.2	µm
			0	mm	500	mm	Exy diagonal axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	3.2	µm
			0	mm	900	mm	Exy diagonal axis/ (20 ± 2) °C / (resolution: 0.0001 mm)	3.8	µm
			0	mm	50	mm	X Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	2.9	µm
			0	mm	300	mm	X Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	2.8	µm
			0	mm	500	mm	X Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	3.8	µm
			0	mm	780	mm	X Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	5.5	µm
			0	mm	50	mm	Y Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	1.8	µm
			0	mm	300	mm	Y Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	2.5	µm
			0	mm	500	mm	Y Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	3.6	µm
			0	mm	580	mm	Y Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	4.1	µm
			0	mm	150	mm	Z Axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	3.5	µm
			0	mm	50	mm	Exy diagonal axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	3.3	µm
			0	mm	300	mm	Exy diagonal axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	3.2	µm
			0	mm	500	mm	Exy diagonal axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	4.0	µm



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA4010 video measurement system (linearscale type) (on-site calibration)	Mitutoyo /BM1-8M-0 Nikon/50 mm Nikon/300 mm Mitutoyo /182-532 Mitutoyo /HL3-1000	In-house method: Calibration Procedure for video measurement system (on-site calibration) (Document No.: GCH-SCP-L07)	0	mm	900	mm	Exy diagonal axis/ (20 ± 5) °C / (resolution: 0.0001 mm)	6.3	μm
			0	mm	50	mm	X Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	2.6	μm
			0	mm	300	mm	X Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	4.2	μm
			0	mm	500	mm	X Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	6.6	μm
			0	mm	780	mm	X Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	11	μm
			0	mm	50	mm	Y Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	1.8	μm
			0	mm	300	mm	Y Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	4.0	μm
			0	mm	500	mm	Y Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	6.5	μm
			0	mm	580	mm	Y Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	7.5	μm
			0	mm	150	mm	Z Axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	6.5	μm
			0	mm	50	mm	Exy diagonal axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	3.4	μm
			0	mm	300	mm	Exy diagonal axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	4.4	μm
			0	mm	500	mm	Exy diagonal axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	6.7	μm
			0	mm	900	mm	Exy diagonal axis/ (20 ± 10) °C / (resolution: 0.0001 mm)	12	μm
Approval Signatory: HUNG, Chuan-Hsi; CHEN, Chih-Chuan									

Note : Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.
(Null Below)

