



財團法人全國認證基金會

Taiwan Accreditation Foundation

Certification Accreditation

(Certificate No : L0956-220512)

This is to certify that

GONG-Hung Technology Co., Ltd.

GCH Technology Calibration Laboratory(Temperature)

No.1, Ln. 36, Wenfeng St., Fengshan Dist., Kaohsiung City 830, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2017 ; CNS 17025:2018

Accreditation Number : 0956

Originally Accredited : August 01, 2003

Effective Period : July 19, 2022 to July 18, 2025

Accredited Scope : Calibration Field, see described in the Appendix



Scan to verify

Ching-Chang Lien

Ching-Chang Lien

President, Taiwan Accreditation Foundation

May 12, 2022

Accreditation Number : 0956

Laboratory Head : HUNG, Chuan-Hsi

Temperature/Humidity

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KE1001 Liquid-in-glass Thermometer (Total immersion)	SHINNIHON KEISOKU/48~102 °C SHINNIHON KEISOKU/98~152 °C SHINNIHON KEISOKU/148~202 °C	liquid-in-glass thermometer calibration procedure (Document No.: GCH-SCP-T01)	60	°C	200	°C		0.5	°C
Approval Signatory: HUNG, Chuan-Hsi									
KE1006 Temperature Indicator (on-site calibration included)	FLUKE/754	Thermometer (on-site calibration included) (Document No.: GCH-SCP-T03)	-100	°C	1000	°C	Type E (on-site calibration included)	0.9	°C
			-100	°C	1300	°C	Type N (on-site calibration included)	1.4	°C
			-100	°C	1200	°C	Type J (on-site calibration included)	1.1	°C
			-100	°C	1350	°C	Type K (on-site calibration included)	1.1	°C

The Appendix forms an integral part of this Certificate, which shall be invalid when use without the Appendix

P2, total 7 pages



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KE1006 Temperature Indicator (on-site calibration included)	FLUKE/754	Thermometer (on-site calibration included) (Document No.: GCH-SCP-T03)	-200	°C	400	°C	Type T (on-site calibration included)	1.0	°C
			0	°C	1750	°C	Type R (on-site calibration included)	2.0	°C
			0	°C	1750	°C	Type S (on-site calibration included)	2.3	°C
			-100	°C	800	°C	PT100 (on-site calibration included)	0.7	°C
Approval Signatory: HUNG, Chuan-Hsi									
KE1007 Radiation Thermometer	Radiation Thermometer FLUKE/566	Calibration procedure of the radiation thermometers (Document No.: GCH-SCP-T02)	-20	°C	50	°C	emissivity1.00 (include) below	3.4	°C
			50	°C	500	°C	emissivity0.95 (include) below	3.8	°C
Approval Signatory: HUNG, Chuan-Hsi									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KE1008 Thermocouple Simulator (on-site calibration included)	FLUKE/754	Thermometer (on-site calibration included) (Document No.: GCH-SCP-T03)	-100	°C	1000	°C	Type E (on-site calibration included) (IN)	0.9	°C
			-100	°C	1300	°C	Type N (on-site calibration included) (IN)	1.4	°C
			-100	°C	1200	°C	Type J (on-site calibration included) (IN)	1.1	°C
			-100	°C	1350	°C	Type K (on-site calibration included) (IN)	1.1	°C
			-200	°C	400	°C	Type T (on-site calibration included) (IN)	1.0	°C
			0	°C	1750	°C	Type R (on-site calibration included) (IN)	2.0	°C
			0	°C	1750	°C	Type S (on-site calibration included) (IN)	2.3	°C
			-100	°C	800	°C	PT100 (on-site calibration included) (IN)	0.7	°C
			-100	°C	1000	°C	Type E (on-site calibration included) (OUT)	1.1	°C
			-100	°C	1300	°C	Type N (on-site calibration included) (OUT)	1.4	°C
			-100	°C	1200	°C	Type J (on-site calibration included) (OUT)	1.1	°C
			-100	°C	1350	°C	Type K (on-site calibration included) (OUT)	1.1	°C



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KE1008 Thermocouple Simulator (on-site calibration included)	FLUKE/754	Thermometer (on-site calibration included) (Document No.: GCH-SCP-T03)	-200	°C	400	°C	Type T (on-site calibration included) (OUT)	1.0	°C
			0	°C	1750	°C	Type R (on-site calibration included) (OUT)	2.0	°C
			0	°C	1750	°C	Type S (on-site calibration included) (OUT)	2.3	°C
			-100	°C	800	°C	PT100 (on-site calibration included) (OUT)	0.7	°C
Approval Signatory: HUNG, Chuan-Hsi									
KE1009 Temperature Calibrator (on-site calibration included)	FLUKE/754	Thermometer (on-site calibration included) (Document No.: GCH-SCP-T03)	-100	°C	1000	°C	Type E (on-site calibration included) (IN)	0.9	°C
			-100	°C	1300	°C	Type N (on-site calibration included) (IN)	1.4	°C
			-100	°C	1200	°C	Type J (on-site calibration included) (IN)	1.1	°C
			-100	°C	1350	°C	Type K (on-site calibration included) (IN)	1.1	°C
			-200	°C	400	°C	Type T (on-site calibration included) (IN)	1.0	°C
			0	°C	1750	°C	Type R (on-site calibration included) (IN)	2.0	°C
			0	°C	1750	°C	Type S (on-site calibration included) (IN)	2.3	°C



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KE1009 Temperature Calibrator (on-site calibration included)	FLUKE/754	Thermometer (on-site calibration included) (Document No.: GCH-SCP-T03)	-100	°C	800	°C	PT100 (on-site calibration included) (IN)	0.7	°C
			-100	°C	1000	°C	Type E (on-site calibration included) (OUT)	1.1	°C
			-100	°C	1300	°C	Type N (on-site calibration included) (OUT)	1.4	°C
			-100	°C	1200	°C	Type J (on-site calibration included) (OUT)	1.1	°C
			-100	°C	1350	°C	Type K (on-site calibration included) (OUT)	1.1	°C
			-200	°C	400	°C	Type T (on-site calibration included) (OUT)	1.0	°C
			0	°C	1750	°C	Type R (on-site calibration included) (OUT)	2.0	°C
			0	°C	1750	°C	Type S (on-site calibration included) (OUT)	2.3	°C
			-100	°C	800	°C	PT100 (on-site calibration included) (OUT)	0.7	°C
Approval Signatory: HUNG, Chuan-Hsi									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KE1011 Sensor /Indicator of Temperature Controlled Chamber (on-site calibration included)	FLUKE/754	Thermometer (on-site calibration included) (Document No.: GCH-SCP-T03)	-100	°C	1000	°C	Type E (on-site calibration included) (IN)	0.9	°C
			-100	°C	1300	°C	Type N (on-site calibration included) (IN)	1.4	°C
			-100	°C	1200	°C	Type J (on-site calibration included) (IN)	1.1	°C
			-100	°C	1350	°C	Type K (on-site calibration included) (IN)	1.1	°C
			-200	°C	400	°C	Type T (on-site calibration included) (IN)	1.0	°C
			0	°C	1750	°C	Type R (on-site calibration included) (IN)	2.0	°C
			0	°C	1750	°C	Type S (on-site calibration included) (IN)	2.3	°C
			-100	°C	800	°C	PT100 (on-site calibration included) (IN)	0.7	°C
Approval Signatory: HUNG, Chuan-Hsi									

Note: Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.

(Null Below)

